

Certificate of Analysis

Product Name	PVDF, 0.45μm, 265mm*3750mm		
Cat.No.	PF0045	Storage Conditions	Sealed storage at room temperature to avoid moisture
Lot.No.	251219E01-01	Expiration Date	18-12-2030
Molecular Formula	N/A	Molecular Weight	N/A
CAS Number	N/A	Excitation Emission	N/A
Quality Control			
Items	Specification	Result	
Packaging appearance	Intact and undamaged	Pass	
Appearance	The surface of the filter material should be clean, flat, intact, without mechanical damage, and without fiber shedding; The sheet should have no obvious burrs or brushed edges.	Pass	
Bubble-point pressure	The alcohol bubble-point pressure shall be within the range of 0.035-0.1 MPa.	Pass	
Water flow rate	The water flow rate shall be no less than 12,000 LMH/bar	Pass	
Thickness	0.10-0.18mm	Pass	
Appendix			
This Product has been tested and the components have been found to conform to UE quality standards.			



Conclusion	Up to standard	Report Date	09-06-2026
QC	<i>Cen ming Jiang</i>	QA	<i>Qin Hua-Yu</i>

